

Attachment 10

DATA ITEM DESCRIPTION			Form Approved OMB No 0704-0188	
1 TITLE Special Equipment Tools and Test Equipment List		2 IDENTIFICATION NUMBER DI-ILSS-80868		
3 DESCRIPTION / PURPOSE 3.1 A listing of unique tools and test equipment required for the repair and use of the end item. The listing is intended to be used by the government to determine provisioning requirements for special equipment.				
4 APPROVAL DATE (YYMMDD) 890829	5 OFFICE OF PRIMARY RESPONSIBILITY (OPR) S/DPSC-RST	6a DTIC APPLICABLE	6b GIDEP APPL. CABLE	
7 APPLICATION / INTERRELATIONSHIP 7.1 This DID contains the format and content preparation instructions for the data product generated by the specific and discrete task requirement as delineated in the contract.				
8 APPROVAL LIMITATION		9a APPLICABLE FORMS	9b AMSC NUMBER S4771	
10 PREPARATION INSTRUCTIONS <p>10.1 <u>General</u>. The list shall contain a listing of unique items required to inspect, test, calibrate, service, repair or overhaul the end item. The list should be comprised of items other than those listed in Table I.</p> <p>10.2 <u>Format and Content</u>. The list shall be in the format shown on the sample data sheet in Table II. The first page of the listing shall include the identification information described in 10.2.1. The listing data shall have the format and content described in 10.2.2 to 10.2.6. Note, sample data is included in Table II for clarity.</p> <p>10.2.1 <u>Heading Data</u>. The top portion of the list shall consist of information to permit the identification of the end item with the listing i.e. Item Name, NSN, Contractor Name (Source), Reference or Part number, and Contract number.</p> <p>10.2.2 <u>Line Item Control</u>. Column 1 of the list shall be used for sequential line item control, commencing with the first line item (Item No.) on the first page and continuing to the last line item on the last page of the listing. Each line item shall represent a single support item.</p> <p style="text-align: right;">Continued on Page 2</p>				
11 DISTRIBUTION STATEMENT <u>DISTRIBUTION STATEMENT A</u> : Approved for public release; distribution is unlimited.				

Block 10, Preparation Instructions (Continued)

10.2.3 Commercial and Government Entity code(s). Column 2 of the list shall indicate the Commercial and Government Entity (CAGE) code(s) for each line item. Each line item shall list one or more entries. The first entry shall reflect the original manufacturers code (if available). Additional entries shall be included for the end item manufacturer or other sources for the support item.

10.2.4 Catalog Number. Column 3 of the list shall indicate the part or catalog number assigned to the support item for each CAGE. If no CAGE code exists the part number shall still be listed. Refer to 10.2.3 for additional CAGE information.

10.2.5 Approved Item Name. Column 4 of the list shall indicate the Approved Item Name which is listed in the H6 series of the Federal Item Name Directory, or the name assigned to the part by the manufacturer.

10.2.6 Unit Price. Column 5 of the list shall indicate an estimate of the Unit Price.

Block 10. Preparation Instructions (Continued)

TABLE I. Test Equipment and Tool Kit List

<u>Nomenclature</u>	<u>KIT</u>	<u>A</u>	<u>B1</u>	<u>B2</u>	<u>C</u>
Calibrator-Analyzer, Pneumatic		X	X	X	X
Defibrillator Tester		X	X	X	X
Electrical Safety Analyzer		X	X	X	X
Electrical Counter		X	X	X	X
Electrosurgical Tester		X	X	X	X
IV Infusion Pump Analyzer				X	X
Light Meter		X	X	X	X
Multimeter, Digital		X	X	X	X
Multimeter, Digital, mas			X	X	X
Oscilloscope, Dual Trace, Storage			X	X	X
Oscilloscope, Storage		X	X	X	X
Patient Simulator, ECG/BP		X	X	X	X
Phototachometer		X	X	X	X
Portable Densitometer		X	X	X	X
Printed Circuit Board Tester				X	X
Shop Set, Depot				X	
Signal Generator		X	X	X	X
Test Set, Electronic Circuit			X	X	X
Tester, Stylus		X			
Tool Kit, Medical Equipment, Repairman		X	X	X	X
Tool Kit, Medical Equipment, Organizational		X	X	X	X
Transistor Tester		X	X	X	X
Ultrasound Radiometer				X	X
X-Ray Calibration and Verification System		X	X	X	X
X-Ray Cassette, KVP		X	X	X	X
X-Ray Pulse Timer		X	X	X	X

Block 10. Preparation Instructions (Continued)

TABLE 11. Sample Data Sheet. Tools and Test Equipment List

End Item Name: <u>Sterilizer, Vacuum</u> End Item Source: <u>Company XYZ</u> Contract No: <u>DLA12087R1234</u>				
End Item NSN: <u>6530-01-123-1234</u> Reference No: <u>Model A-12</u>				
<u>Tools and Test Equipment List</u>				
¹ <u>ITEM No</u>	² <u>CAGE</u>	³ <u>PART No</u>	⁴ <u>ITEM NAME</u>	⁵ <u>UNIT PRICE</u>
0001	12312 80307	E10 E10	Thermometer, Lag with case	94.33
0002	80307	E11A	Indicator, Absolute Pressure, 710B, Heise	973.33
0003	80307	L115	Wrench, Safety Key	18.30
0004	80307	L117	Wrench, Safety Key	6.50
PAGE 1 of 1				